



#### THE STUDY

A comprehensive inventory of the current ARM uncertainty estimates was constructed from inputs by each ARM Instrument Mentor for all instruments and for important variables.

The uncertainty estimate provided was classified as either instrument resolution, precision, accuracy, or none (information not provided or not available).

### THE SPECIFICS

For those cases where uncertainty was reported as accuracy, the method of traceability of the calibration was classified as a traceable standard, consensus procedure, or expert judgment.

Although the results of this study are contained in a spreadsheet, due to the number of instruments and measured variables, there is far too much information to be displayed here.

The results presented here indicate the families of instruments and measured variables that have common attributes of uncertainty type.

## **REPORTING UNCERTAINTIES**

Not all of information about instrument uncertainties is complete, so iterations with the ARM Instrument Mentors will be needed.

Uncertainty estimates can have dependencies on environmental factors and therefore a non-linear component. For example, there is a range dependence on water-vapor mixing ratios from Raman Lidar, which is based on the dryness of the atmosphere.

Displaying instrument measurement uncertainties will be a challenge!



Newsom: ARM Raman Lidar water-vapor mixing ratio error estimates based on comparisons with radiosondes.

Contact the authors: dlsisterson@anl.gov, ecampos@anl.gov



# **A Unified Approach for Reporting ARM Measurement Uncertainty**

**Doug Sisterson, Edwin Campos** Argonne National Laboratory, Environmental Science Division, Argonne, IL 60491



	<b>STUDY</b>	RESULTS		
	Instrument	Measurement	Uncertainty Estimate	Uncertainty Type
	Normal Incidence Pyrheliometer (NIP)	Direct beam normal irradiance	± 3% (>700 Wm <sup>-2</sup> )	Precision: GUM Combined Uncertainty, after calibrations with traceable standard (NIST) and concensus procedure (WISG)
	Multifilter Rotating Shadowband Radiometer <b>(MFRSR)</b>	Clear Skies total horizontal irradiance	± 2.1 %	Accuracy: Calibrations using traceable standards (NIST) and consensus procedure (Langley plots)
	<section-header></section-header>	Reflectivity	4 dB	Other: Combination of calibration of components, literature, and expert opinion. Calibration is highly idealized, assumes no atmospheric losses, a known target in the far field, the return is from the target only, and no multi-path to the target.
BELFORT Model AEPG II 600/1000	Rain Gauge – Belfort Model AEPG 600 Weighing Bucket	Rainfall amount (accumulation)	+/- 0.25mm (0.01 inches)	Resolution (minimum detectable signal)

Corresponds to instrument calibrations Accuracy to address known measurement errors. Corresponds to minimum detectable Resolution signals of measurement. Corresponds to the variance from Precision repeated calibrated measurements under field conditions. Corresponds to theoretical and/or Other empirical arguments and/or expert judgment Corresponds to measurements of None unknown uncertainty (no estimates were provided).





